



www.free-foundation.org



bedside commode

A qualified medical professional should educate and train users for appropriate use. It is your responsibility to use this device correctly.

Commodes can be used at the bedside or in the bathroom over the toilet. All commodes have adjustable heights using external leg extensions and are adjusted according to the patient's height. Make sure all legs are set at the same height. The legs also have non-skid rubber tips for increased stability and safety. Commodes are available with armrests for added comfort and to assist the patient when lifting off the commode.

It is important to remember that commodes have a base weight limit of 300 lbs.

A qualified medical professional should install bedside commode to ensure appropriate installation. The rubber tips should be inspected regularly. Worn or damaged tips should be replaced immediately.

It is your responsibility to maintain and provide upkeep of this item. It is your responsibility to use this device correctly.