



www.free-foundation.org

Worn or damaged tips should be replaced immediately.

hemi-walker

A qualified medical professional should educate and train users for appropriate use. It is your responsibility to use this device correctly.

The side style hemi-walker is designed for individuals with the use of only one hand or arm. It is also ideal for individuals who do not require a walker but need a wider base for support. This product is lighter than a walker and more stable than a cane, ensuring user safety. The rubber tips should be inspected regularly.

## **Height Adjustment**

The height of the hand grips should be such that is provides a slight bend in the elbow when you are standing straight and holding the hand grip. A 20-to-30degree bend in the elbow is usually considered desirable. A qualified medical professional should adjust the height of all four walker legs to place the height of the hand grips slightly above your wrists when in a standing position. After making the height adjustment, check carefully to ensure that all four legs are locked securely in the selected position.

It is your responsibility to maintain and provide upkeep of this item. It is your responsibility to use this device correctly.